



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell
Single-Ended Shear Beam, Compression
Model: SE-CG Series (see page 2)
 n_{\max} : Multiple Cell 5 000
 v_{\min} : (see page 2)
Capacity: 250 lb to 20 000 lb
Accuracy Class: III

***Submitted By: Contact Info. Updated: December 09**

Coti Global Sensors, Inc.
5709 Highway 53
Harvest, AL 35749
Tel: 256-852-9900
Fax: 256-852-9903
Contact: Amy Allen
Email: amy@cotiglobal.com
Web site: www.cotiglobal.com

Standard Features and Options**Standard Features:**

- Alloy Steel or Stainless Steel Construction
- Method of Sealing: Potted with Metal Cover
- Number of Wires: 4 wires
- Excitation Voltage: 10 VDC
- Nominal Output: 3.0 mV/V or 2.0 mV/V
- Bridge Resistance Input Nominal: 350 Ohms

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Randy Jennings
Chairman, NCWM, Inc.

Judy Cardin
Chairman, National Type Evaluation Program Committee
Issued: December 22, 2009

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Coti Global Sensors, Inc.

Load Cell / SE-CG Series

Application: The load cells may be used in Class III scales for multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this Certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the v_{\min} values, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions (n_{\max}) and with larger v_{\min} values than those listed on the Certificate. However, the load cells must be marked with the appropriate n_{\max} and v_{\min} for which the load cell may be used.

Identification: A pressure sensitive, tamper evident, identification badge containing the manufacturer, model designation, and serial number is located on the load cell. All other required information, if not marked on the load cell, must be on an accompanying document including the serial number of the load cell.

Models	Capacity (lb)	v_{\min} (lb)	Minimum Dead Load (lb)
SE-CG-23	250	0.038	10
SE-CG-23-SS	500*	0.075	10
SE-CG-23-LP	1000	0.14	10
SE-CG-23-SSW	1250	0.175	10
SE-CG-SB3	1500	0.21	10
SE-CG-MK15	2000	0.28	10
SE-CG-SB250	2500	0.35	50
SE-CG-SSB	3000	0.42	50
SE-CG-743	4000	0.56	88
SE-CG-745	5000*	0.70	88
	10 000	1.40	88
	15 000	2.10	88
	20 000	2.80	88

* Two Load Cells Tested

Test Conditions: Two 500-lb capacity and two 5000-lb. load cells were tested at NIST using dead weights as the reference standard. The data were analyzed for multiple load cell applications. The cells were tested over a temperature range of -10 °C to 40 °C. Three tests were run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test was waived due to the insensitivity of the load cell design to changes in barometric pressure.

Evaluated By: NIST Force Group, NIST Office of Weights and Measures

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2004. NCWM, Publication 14: Weighing Devices, 2004.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: S. Patoray, L. Bernetich (NCWM) 04-066; J. Truex (NCWM) 08-074

Examples of Device:

